

Certificate of Traceability and Calibration

13.5% Gravimetric Sucrose Standard

Lot Number: SC110818A

Accuracy: $\pm 0.1\%$ (if stored correctly)



1295 Morningside Avenue, Unit 16-17
Toronto, ON M1B 4Z4 Canada
Phone: 416-261-4865 Fax: 416-261-7879
www.scigiene.com

Company: INSERT COMPNAY NAME

Date Prepared: November 8th, 2018

Expiration Date: November 8th, 2019

This product was Calibrated and Certified to be traceable to National Institute of Standards and Technology standards as per the following validation information.

TEST EQUIPMENT USED BY SCIGIENE FOR PREPARATION:

Acculab VIC-212 Top Loading Scale, Serial#: 167555, Repeatability $\pm 0.01\text{g}$ *Certified with:*

Mettler-Toledo AG Single weight 5g, OIML F1, Serial Number: B236291931, Tolerance $\pm 0.16\text{mg}$

Mettler-Toledo AG Single weight 20g, OIML F1, Serial Number: B236291932, Tolerance $\pm 0.25\text{mg}$

Volumetric Flask, 100ml $\pm 0.08\text{ml}$ at 20°C

PREPARED USING:

ACS grade Sucrose (Lot #: 0316C505) Exp Date: 10/2020

Distilled Water

Technician: Harvi Ancheta

A handwritten signature in dark ink, appearing to read "Harvi Ancheta", is written over the printed name.