

Certificate of Traceability and Calibration



1295 Morningside Avenue, Unit 16-17
Toronto, ON M1B 4Z4 Canada
Phone: 416-261-4865 Fax: 416-261-7879
www.scigiene.com

Company: **INSERT COMPANY NAME**
Certificate Number: **SC11.27.18C**
Certification Date: **November 27th, 2018**
Expiration Date: **November 27th, 2019**

This product was Calibrated and Certified to be traceable to National Institute of

Standards and Technology standards as per the following validation information.

Test equipment used by for verifying specifications and accuracy during calibration procedures:

Reference Standards

Milwaukee pH 7.01 Buffer, MA9007, Accuracy; pH = 7.01 ± 0.01 @25C
Lot 0074 Exp: 04/2021

Milwaukee pH 4.01 Buffer, MA9004, Accuracy; pH = 4.01 ± 0.01 @25C
Lot 9848 Exp: 02/2021

Model #	Serial #	Reference Standard(s)	Instrument Reading As Found	Instrument As Left	Difference As Left	Specification Acceptance
Hanna HI 99161 pH Meter with FC202 probe SN. 0343005I	E0007783	4.01 pH	4.04 pH	4.01 pH	0.00 pH	± 0.02 pH
		7.01 pH	7.05 pH	7.00 pH	0.01 pH	± 0.02 pH

Technician: Harvi Ancheta

A handwritten signature in black ink, appearing to read "Harvi Ancheta", is written over the printed name.